

proFIRE  
User Manual

# 5x Buffer B

running buffer for proFIRE

Dynamic Biosensors GmbH & Inc.  
PF-BU-B-5 v5.1



## Product Description

Order Number: **PF-BU-B-5**

Table 1. Contents and Storage Information

Material	Composition	Amount	Storage
<b>5x Buffer B</b>	250 mM Na <sub>2</sub> HPO <sub>4</sub> /NaH <sub>2</sub> PO <sub>4</sub> , 5 M NaCl, pH 7.2; 0.2 µm sterile filtered	50 mL	RT

For research use only.

This product has a limited shelf life, please see expiry date on label.

At temperatures lower than 18°C precipitation may occur.

## Preparation

Dilute the complete solution 5x Buffer B pH 7.2 (50 mL) by mixing with 200 mL ultrapure water.

After dilution Buffer B is ready for use (50 mM Na<sub>2</sub>HPO<sub>4</sub>/NaH<sub>2</sub>PO<sub>4</sub>, 1 M NaCl).

The diluted buffer should be stored at 2-8°C.

## Contact

**Dynamic Biosensors GmbH**

Perchtinger Str. 8/10  
81379 Munich  
Germany

**Dynamic Biosensors, Inc.**

300 Trade Center, Suite 1400  
Woburn, MA 01801  
USA

**Order Information** [order@dynamic-biosensors.com](mailto:order@dynamic-biosensors.com)

**Technical Support** [support@dynamic-biosensors.com](mailto:support@dynamic-biosensors.com)

[www.dynamic-biosensors.com](http://www.dynamic-biosensors.com)

Instruments and chips are engineered and manufactured in Germany.

©2024 Dynamic Biosensors GmbH | Dynamic Biosensors, Inc. All rights reserved.