


switchSENSE® Biochip MPC-48-2-R1-S

Product Description


Product Code	MPC-48-2-R1-S
Suitable for	switchSENSE® DRX ^{red} & DRX ²
Storage	Store at 2-8 °C, dry in the dark
Layout	2 reference spots (nanolever A) and 4 probe spots (nanolever B)
Grade	Standard

Multi-purpose chips are versatile tools to analyze protein-protein, protein-small molecule, DNA/RNA-protein and DNA/RNA-small molecule interactions.


Biochip Design




Biochip with 4 flow channels, each with multiple detection spots in series



6 detection spots in 1 flow channel



Each detection spot: nanolevers with red fluorescent dyes




Nanolevers with the same sequence and fluorescent dye

The detection spots are functionalized with one nanolever, either NL-B48 on probe spots or NL-A48 on reference spots. The 5'-end of the DNA is fixed on the surface while the 3'-end carries a red fluorescent dye (R1).

Biochip Layout | Flow channel with 2 reference (A) & 4 probe (B) spots (end-point referencing in same channel)

This biochip - presenting 2 reference sequence (NL-A48) and 4 probe sequence (NL-B48) spots - enables to measure the reference in the same flow channel.



For more information about the DNA sequences, please contact us at info@dynamic-biosensors.com.

DNA for the Biochip

Each Biochip comes with unmodified complementary DNA:

cNL-A48 (500 µL, c = 1 µM)

cNL-B48 (500 µL, c = 1 µM)

Application Areas

Binding Kinetics	k_{on} , k_{off} , K_D
Binding Affinity	K_D , titration curve, n (Hill coefficient)
Protein Diameter	D_H (Hydrodynamic diameter)
Conformational Change	ΔD_H (relative change of hydrodynamic diameter)
Melting & Thermodynamics	T_M , ΔG , ΔH , ΔS
Multimers & Aggregation	Monomer-dimer discrimination, aggregation
Nuclease & Polymerase Activity*	k_{cat} , K_M , $T_{activate}$
Bispecific Binders & Avidity*	Binding affinity/avidity/kinetics;

* for this application, other biochips (ENZ or BIF) are recommended

Biochip Handling

The biochip is ready to use. For research only.

Avoid touching or picking up the biochip with your hands as this may cause biolayer electrostatic discharge. Instead, always use the vacuum tweezers provided with the instrument when handling the biochips.

After installing the biochip in the DRX instrument and before starting an experiment be sure that the selected channel is passivated with passivation solution (Order No. SOL-PAS-1-5).

More information about biochip storage and handling can be found in the document "switchSENSE Biochips Storage Information".

Grade I Standard

Standard grade chips are batch tested which means one out of a batch undergoes a comprehensive quality control – proper switching dynamics, calibration (voltage-response of the DNA layer) and fluorescence levels. In addition each biochip is tested for proper fluorescence levels on each electrode. Due to the production process, not all detection spots fulfill our high quality standards. Each standard grade biochip is certified as having 20 or more active detection spots.

Compatible Functionalization Kits

All compatible functionalization kits for this biochip are listed on our website:

www.dynamic-biosensors.com/reagents/

Contact

Dynamic Biosensors GmbH
 Lochhamerstr. 15
 82152 Martinsried/Planegg
 Germany

Order Information

Phone: +49 89 89 74 544 0
 order@dynamic-biosensors.com

Dynamic Biosensors Inc.
 9705 Carroll Centre Road, Suite 100
 San Diego, CA 92126
 USA

Technical Support

Phone: +49 89 89 74 544 66
 support@dynamic-biosensors.com